

FAA Compliance for ResMed Germany Inc. Devices

To whom it may concern,


The US Department of Transportation (DOT) Final Rule, "Nondiscrimination on the Basis of Disability in Air Travel" (73 FR 27614 which updates Title 14 CFR Part 382), effective May 13, 2009 provides important new requirements for the accommodation of passengers with respiratory assistive devices (Ventilators, Respirators and CPAP machines).

In line with these requirements, respiratory assistive devices may be used onboard an aircraft, without further testing by the carrier, provided they have been tested for Electromagnetic Compatibility (EMC) in accordance with the current version of RTCA/DO-160, Section 21, Category M.

ResMed Germany Inc. has now successfully completed testing for the following respiratory assistive devices. **The devices listed below comply with RTCA/DO-160G, Section 21, Category M, and are considered FAA compliant.**

Product Code	Product Description
24140, 24141, 24142, 24143, 24144, 24145, 24146, 24147, 24160	Stellar 150
24151, 24152, 24153, 24154, 24155, 24156, 24157, 24158, 24159	Stellar 100

Kind regards,


Sandra Grünwald
Manager – Quality Assurance & Regulatory Affairs
ResMed Germany Inc.

22. November 2011